

PROPER CARE AND MAINTENANCE
OF SILVER PLATED
SURGICAL INSTRUMENTS

The M14-0500 Sims Uterine Sound is made from Copper Wire which is then Silver Plated. Due to Oxidation the color will change.

The instrument can be cleaned by using SILVO paste and rubbing it with a soft cloth.

They can be sterilized either by gas or by cold sterilization. Sterilizing silver plated instruments in heat or steam will cause them to rust.